



CRB Analyse Service GmbH
Stefan Pierdzig
Bahnhofstrasse 14
D-37181 Hardegsen - Germany

VSL
Thijsseweg 11
2629 JA Delft
P.O. Box 654
2600 AR Delft
The Netherlands
T 015 269 15 00
F 015 261 29 71
E info@vsl.nl
I www.vsl.nl

Your reference

Our reference
RIAB 2020-83

Your reference

subject
RIAB 83 results

Dear participant,

Please find enclosed the evaluation report of RIAB 83.

If you have any questions, please don't hesitate to contact us.

Regards

VSL PT-Staff
Peter van Otterloo

Description	RIAB - Ring Investigation Identification Asbestos in Bulk Materials.
Participant	CRB Analyse Service GmbH Bahnhofstrasse 14 37181 Hardegsen Germany
Method and evaluation	The PT was organised in accordance with document: Protocol RIAB. After obtaining the reference value, the samples have been shipped to the participants laboratory. The reported results of the participant are compared with the reference values. The results of the comparison are presented on page 2.
Policy statement:	The evaluation reports of VSL's proficiency testing schemes are provided for the purpose of communicating the proficiency demonstrated by participants on specific measurements and/or calibrations. The reports are intended to be used in support of demonstrating competence in measurement or calibration, fulfilling quality control requirements as stipulated in written standards on showing such competence, and claims of calibration and measurement capabilities.

Delft, 28 October 2020
VSL B.V.

R.P. van Otterloo
Project manager comparisons



Period of ILC August 2020 – October 2020


Results &
Evaluation

Fibre type	8301		8302	
	Results	Reference	Results	Reference
Chrysotile	< 0,1%	< 0,1%	10-15%	10-15%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
	Points: 10		Points: 10	

Fibre type	8303		8304	
	Results	Reference	Results	Reference
Chrysotile	> 60%	30-60%	2-5%	2-5%
Amosite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Crocidolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Tremolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Actinolite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
Anthophyllite	< 0,1%	< 0,1%	< 0,1%	< 0,1%
	Points: 10		Points: 10	

Results The result with a total score of 100 is satisfactory

Result over RIAB 80-83 The result of RIAB 80 to 83 with a total score of 95 is satisfactory



RIAB 83 results

